

# Attestation of Conformity

No. E8A 125010 0001 Rev. 01

**Holder of Attestation:** **TOSUN Limited**  
Unit 417, 4/F Lippo CTR Tower Two, No.89 Queensway  
00852 Admiralty  
HONG KONG

**Name of Object:** **Electrical equ. for measurement, control  
and laboratory use  
4 channels CANFD interface**

**Model(s):** **TC1014**


**Description of  
Object:** Rated voltage: 5Vd.c.  
Rated current: 0.44A

**Tested  
according to:** EN IEC 61326-1:2021

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for conformity assessment. For details see: [www.tuvsud.com/ps-cert](http://www.tuvsud.com/ps-cert)

**Test report no.:** 708882309187-01

**Date,** 2024-07-05

  
( Linda Qiu )